



## CERTIFICATO DI ANALISI - TEST CERTIFICATE

LOTTO/LOT	ARTICOLO/ARTICLE	C.V (limit 10 %)	data/date
<b>M0993</b>	<b>MT0F2-MB</b>	<b>&lt; 10 %</b>	<b>29/4/14</b>

Si certifica che il lotto sopra indicato è stato prodotto secondo le procedure operative standard.

Questo prodotto è stato analizzato ed è risultato conforme alle specifiche stabilite.

### Metodo di analisi

Soluzione di coating: 5 µg/ml di Rabbit IgG in 0.1M Tampone Carbonato pH 9.6.  
Coniugato : Goat Anti-IgG-HRP

### Controllo finale

Il prodotto finito è stato ispezionato ed approvato per le seguenti caratteristiche fisiche:

Test	Risultato
esame visivo	approvato
aspetto generale	approvato
imballaggio	approvato

Biomat hereby certifies that this lot has been manufactured according to applicable standard operating procedures.

This product has been tested and has passed all established specifications

### Method of analysis

Coating solution: 5 µg/ml of Rabbit IgG in 0.1 M Carbonate Buffer pH 9.6.  
Conjugate: Goat Anti-IgG-HRP

### Final control

This product has been tested and approved for the following physical characteristics:

Test	Result
visual examination	approved
general appearance	approved
packaging	approved

**Dr. Maurizia Pettenati**  
Quality Control manager

*Copia conforme al certificato originale firmato e conservato presso i nostri archivi  
Original signed certificate is filed in our offices*