

CERTIFICATO DI ANALISI - CERTIFICATE OF ANALYSIS

LOTTO/LOT	ARTICOLO/ARTICLE	data/date	Expiry date
LYS1576	MT12F2-LYS-L	13/06/17	06/2019

Biomat hereby certifies that this lot has been manufactured according to applicable standard operating procedures, current ISO 9001:2008 regulation.

This product has been tested and has passed all established specifications

TEST PROTOCOL

This test measures the interaction of Biotin ester (NHS-biotin) with the ε- amino groups on Lysine available for multipurposes interactions

Method of analysis

Coating solution: 12.5 - 0 µg/ml NHS-biotin solutions diluted in 0.1M PBS- Tween[®] 20 0.15% pH 7.2

Conjugate: 100 ng/ml streptavidin-HRP

value	Specification	Lot results
OD at 0	< 0.25	0.071
OD at 12.5	> 1.500	1.886

Dr. Maurizia Pettenati
 Quality Control manager

Copia conforme al certificato originale firmato e conservato presso i nostri archivi
Original signed certificate is filed in our offices