



CERTIFICATO DI ANALISI - CERTIFICATE OF ANALYSIS

LOTTO/LOT	ARTICOLO/ARTICLE	data/date	Expiry date
LYS1711	MT12F2-LYS-L	20/04/18	04/2020

Biomat hereby certifies that this lot has been manufactured according to applicable standard operating procedures, current ISO 9001:2015 regulation.

This product has been tested and has passed all established specifications

TEST PROTOCOL

This test measures the interaction of Biotin ester (NHS-biotin) with the ε- amino groups on Lysine available for multipurposes interactions.

Method of analysis -uniformity

Coating solution: 3.125 µg/ml NHS-biotin solution diluted in 0.1M PBS- Tween[®] 20 0.15% pH 7.2

Conjugate: 50 ng/ml streptavidin-HRP + 1 µg/ml streptavidin

Results: < 10%

Dr. Maurizia Pettenati
 Quality Control manager

Copia conforme al certificato originale firmato e conservato presso i nostri archivi
Original signed certificate is filed in our offices