

## CERTIFICATO DI ANALISI - CERTIFICATE OF ANALYSIS

LOTTO/LOT	ARTICOLO/ARTICLE	data/date	Expiry date
<b>LYS1822</b>	<b>MT12F2-LYS-L</b>	<b>05/11/18</b>	<b>11/2020</b>

Biomat hereby certifies that this lot has been manufactured according to applicable standard operating procedures, current ISO 9001:2015 regulation.

This product has been tested and has passed all established specifications

### TEST PROTOCOL

This test measures the interaction of Biotin ester (NHS-biotin) with the  $\epsilon$ - amino groups on Lysine available for multipurposes interactions.

#### Method of analysis -uniformity

Coating solution: 12,5  $\mu$ g/ml NHS-biotin solution diluted in 0.1M PBS- Tween<sup>®</sup> 20 0.15% pH 7.2

Conjugate: 50 ng/ml streptavidin-HRP

<b>Specification: &lt; 10%</b>	<b>Result: 7,0%</b>
--------------------------------	---------------------

**Dr. Maurizia Pettenati**  
 Quality Control manager

*Copia conforme al certificato originale firmato e conservato presso i nostri archivi  
 Original signed certificate is filed in our offices*