

CERTIFICATO DI ANALISI - TEST CERTIFICATE

LOTTO/LOT	ARTICOLO/ARTICLE	C.V (limit 10 %)	data/date
M02332	MCB0F-MB	< 10 %	12/05/21

Si certifica che il lotto sopra indicato è stato prodotto secondo le procedure operative standard.

Questo prodotto è stato analizzato ed è risultato conforme alle specifiche stabilite.

Metodo di analisi

Soluzione di coating: 2,5 µg/ml di Human IgG in 0.1M Tampone Carbonato pH 9.6.
 Coniugato: Goat Anti-HlgG-HRP

Controllo finale

Il prodotto finito è stato ispezionato ed approvato per le seguenti caratteristiche fisiche:

Test	Risultato
esame visivo	approvato
aspetto generale	approvato
imballaggio	approvato

Biomat hereby certifies that this lot has been manufactured according to applicable standard operating procedures.

This product has been tested and has passed all established specifications

Method of analysis

Coating solution: 2,5 µg/ml of Human IgG in 0.1 M Carbonate Buffer pH 9.6.
 Conjugate: Goat Anti-HlgG-HRP

Final control

This product has been tested and approved for the following physical characteristics:

Test	Result
visual examination	approved
general appearance	approved
packaging	approved

Dr. Maurizia Pettenati

Quality Control manager

Copia conforme al certificato originale firmato e conservato presso i nostri archivi
 Original signed certificate is filed in our offices